


<b>CCFCA</b> <b>2022-004</b>	<b>Standard Operating Guideline</b>	Page 1 of 3
	<p style="text-align: center;"><b>Cumberland County Fire Chief's Association</b></p>	<p><b>REFERENCES:</b> N/A</p> <hr/> <p><i>Forms: N/A</i>  1<sup>st</sup> Draft Reading 06/27/22  2<sup>nd</sup> Dreft Reading 0725/22  3<sup>rd</sup> Final Reading 08/22/22/</p>
<b>G5-Defibrillator AED ICPR Pads Replacement / Exchange</b>		
APPROVED BY: <i>Freddy L. Johnson Sr.</i> President – <u>Administrative SOG</u> <i>Freddy L. Johnson Sr.</i>		EFFECTIVE: <b>September 1, 2022</b>

**1.0 SCOPE**

This guideline shall apply to all members of the Cumberland County Fire Chief’s Association (CCFCA) and shall be adhered to by all members.

**2.0 PURPOSE**

The purpose of this guideline is to provide a systematic procedure for the distribution, exchange, and rotation of AED Pads for all CCFCA partner agencies. The intent is to limit waste and utilize AED Pads within system before AED Pads reach their expiration date.

**3.0 DEFINITIONS**

**2742-26002 / G5 ICPR AED Pad:** AED Defibrillation Pads - ICPR Pads. The non-polarized adult pads with the Intellisense® CPR (ICPR) Feedback device for use with Powerheart® G5 AED delivers defibrillation therapy and CPR feedback in accordance with the 2015 AHA Resuscitation Guidelines.

**2742-40002 / Cardiac Science G5 Adult** Electrodes Intellisense Non-Feedback

**3.1.1 Standard Operating Guidelines (SOG)** - This SOG specifies the process to request replacement AED Pads and rotating AED Pads with expiring manufactured dates.

**1.1.1 Responsibilities**

1.1.1.1 Cumberland County Fire Departments

1.1.1.2 The President of the CCFCA assigned the responsibility to monitor and control the issue of partner agency AED Pads to the City of Fayetteville Fire Department (FFD). The COF Fire Chief further delineated that responsibility to the FFD EMS Chief.

1.1.1.3 Any Cumberland County Fire Department that needs a replacement set of AED Pads shall make a written request through email to the assigned AED Coordinator, EMS Division Chief with the Fayetteville Fire Department. ([kevinmorgan@fayettevillenc.gov](mailto:kevinmorgan@fayettevillenc.gov)).

- 1.1.1.4 The written request for replacement AED Pads should occur a week before the normally scheduled monthly Chiefs Association meeting.
- 1.1.1.5 It is recommended that each Cumberland County Fire Department maintain **four (4)** sets of spare AED Pads at their station.
- 1.1.1.6 It will be the responsibility of each Cumberland County Fire Department to identify any set of AED Pads in their possession **six (6) months** before its manufacture expiration date. Notification will be made to the EMS Division Chief at Fayetteville Fire Department one week before the normally scheduled Chiefs Association meeting. ([kevinmorgan@fayettevillenc.gov](mailto:kevinmorgan@fayettevillenc.gov))
- 1.1.1.7 Other than an unforeseen emergency, AED Pads nearing expiration date and requested replacement AED Pads will be exchanged at the normally monthly scheduled Chiefs Association meeting.
- 1.1.1.8 Any partner agency that fails to identify AED Pads before their manufacture expiration date, will be responsible for replacing that set of pads from within their own department's budget.
- 1.1.1.9 In the event of **an emergency need** for AED pad replacement, partner agencies should call Fayetteville Fire Station 14 @ 910-433-1400 or 910-433-1725 and notify the supervisor on duty that you have an immediate replacement need. The supervisor will have the AED pads ready for you to pick up at your convenience.

## 1.1.2 **Fayetteville Fire Department**

- 1.1.2.1 The EMS Division Chief with the Fayetteville Fire Department will manage the inventory and ordering of AED Pads for all partner agencies.
- 1.1.2.2 Fayetteville Fire Department will maintain an adequate number of AED Pads to support the needs of all.
- 1.1.2.3 Upon receiving a request to provide replacement AED Pads or to exchange AED Pads, the Fayetteville EMS Chief will bring needed AED Pads to the normally scheduled Chiefs Association meeting for delivery to the requesting fire department.

1.1.2.4 AED Pads approaching six months of expiration date will be returned to the AED coordinator for reassignment to high user units and utilized prior to expiration. (Also see Section 1.1.1.6)

**1.1.3 Anticipated Supply Chain Interruptions**

1.1.3.1 Due to current economic and national vendor supply chain interruptions associated with our standard AED G5 Smart Pad, the association has acquired a limited supply of NONE-FEEDBACK or standard Intellisense PADS that will be issued for utilization until the ICPR Pads are readily available for resupply.